## Flex Comfort Sycura<sup>®</sup> Static Antidecubitus Mattress

MATTRESSES AND PILLOWS

# **PERFORMANCE** designed to prevent pressure ulcers



Frequent Usage

Nursing homes Home Care Hospitalization

#### Flex Comfort Sycura®

### **Product description**

Static open-cell polyurethane foam mattress for general, paediatric and neonatal hospital stays. The mattress is equipped with a special, waterproof, breathable, bi-elastic, antistatic cover treated with Zeneca bacteriostatic-fungostatic system.

Indicated for a posture with a low compressive gradient on the musculoskeletal system, useful in incontinent patients and for improved epidemiological control.

CE marked medical device, performance requirements certified UNI 10707 rev. 2003, comfort and anti-decubitus index certification (ergocert D28) and approved in fire reaction class 11M.

Functional capacity minimum Kg 0 maximum Kg 150.

Sag factor - elastic polyurethane comfort index of N 2.76.

Colour ecru.

Product that allows cardiac massage.

Hypoallergenic, non-toxic, odourless and silent product.

Latex Free product.

Radiotranslucent product.

Density 30 kg/m2 and average bearing capacity 3.3 Kpa (Other technical characteristics and densities available on request)

### Available sizes

Product Code	Size	
AC44/3	190x80x14 cm	
AC44/3L	195x85x14 cm	
AC44/3S	150x70x14 cm	for Paediatrics
AC44/3XS	70x40x10 cm	for Neanatology
AC44/3M	190x60x8 cm	for Stretcher

by Synergic Italiana

#### Gardhen Bilance S.r.l.

Via G. Luraghi c/o Consorzio II Sole - lotto S - 80038 Pomigliano d'Arco (NA) - Italia Tel. +39.081.8692160 pbx - Fax +39.081.8692460 E-mail: info@gardhenbilance.it - PEC: gardhenbilance@pec.it Visit our website **www.gardhenbilance.it /** Follow us on (in) (f) (D) (() (P)

Images reproduced in this sheet are for reference only and dimensional values are nominal. It is the manufacturer's right to make changes to the designs without prior notice. Refer to the optional components catalog and/or commercial proposals for identification of standard equipment and selection of any additional components. In case of discrepant data, the contents of the commercial proposals shall prevail. In the event that preferences are not expressed regarding the colors of the devices ordered, the same, will be made in the colors shown in the data sheet/brochure.

For regulatory references, please consult the user manual and declaration of conformity. Dep. EN\_202 rev. 21/10/2024

